

Is **YOUR defibrillator** ready to use in the event of a **Sudden Cardiac Arrest**?

Please complete the below check sheet to ensure your defib is ready for action.

Defib Model Name:
Defib Model Number:
Defib Check Date:
Defib Location:

Defibrillator Checklist	YES	NO	N/A
DEFIB UNIT			
Is your defib in a visible location?			
Is there visible signage at the storage location to easily identify the defib promptly?			
Is your defib in standby mode? (there should be either a green flashing light or Tick or an OK symbol) Please contact our AED specialists if the unit displays an X or is flashing red or the defib is beeping.			
Is your defib registered with Triple Zero (000)?			
DEFIB ACCESSORIES			
Is there one set of Adult Pads with listed expiry dates?			
Is there one set of Child Pads with listed expiry dates?			
Are the items in the AED Prep kit in date and accounted for? Y N N/A 1x razor () 1x shears () 1x set of gloves () 1x face shield () 1x dressing pad () alcohol wipes ()			

Please note: If any accessories such as defibrillator pads or batteries are missing or expired they should be replaced at your earliest convenience to ensure your defibrillator is ready to use in the event of a Sudden Cardiac Arrest.

Other accessories available:

- 416 Stuart Highway, Winnellie NT
- Allchurch Street Alice Springs NT



To obtain a quote or to place an order for any missing or expired items please contact our friendly DEFIB Specialist:

T: 1300 STJOHN **E:** sales@stjohnnt.asn.au

We're here for you

